

# EMC TEST REPORT

No. 1117858-1

## Electromagnetic disturbances

### EQUIPMENT UNDER TEST

Equipment : NeuroFlexor Main Unit  
Type / model : NF.00.10.20  
Manufacturer : Aggero Medtech AB  
Tested by request of : Karolinska Institutet Innovations AB

### SUMMARY

Referring to the performance criteria and the operating mode during the tests specified in this report the equipment complies with the requirements according to the following standards.

EN 60 601-1-2 (2007), Class B, Group 1.

Date of issue: January 12, 2012

Tested by:   
Lars Ahlqvist

Approved by:   
Hasnain Hassanali

Laboratories are accredited by the Swedish Board for Accreditation and Conformity Assessment (SWEDAC) under the terms of Swedish legislation. The accredited laboratory activities meet the requirements in SS-EN ISO/IEC 17 025 (2005). This report may not be reproduced other than in full, except with the prior written approval by Intertek Semko.

Intertek Semko AB

Torshamnsgatan 43, Box 1103, SE-164 22 Kista, Sweden

Telephone +46 8 750 00 00, Fax +46 8 750 60 30

www.intertek.se

Registered in Sweden: No: SE556024059901, Registered office: As address